



7018 Owensmouth Ave. Suite 103  
Canoga Park, CA, 91303  
Phone: 818-710-1281  
Fax: 818-936-0121  
Email: [Info@immunospec.com](mailto:Info@immunospec.com)  
[www.immunospec.com](http://www.immunospec.com)

## PROTEINASE-3 IgG



REF

E22-312

IVD

FOR INVITRO DIAGNOSTIC USE ONLY

### INTENDED USE

The ImmunoSpec Inc. Proteinase-3 IgG ELISA test system is intended for the qualitative and semi-quantitative detection of anti-Proteinase-3 IgG antibody in human serum. The test system is intended to be used as an aid in the diagnosis of various autoimmune vasculitic disorders characterized by elevated levels of anti-neutrophil cytoplasmic antibodies (ANCA). PR-3-ANCA may be associated with autoimmune disorders such as Wegener's granulomatosis, ICGN, MPA, and PRS. This test is for in vitro diagnostic use.

### SIGNIFICANCE AND BACKGROUND

Anti-neutrophil cytoplasmic antibody (ANCA) was initially described by Davies, et al in 1982 (1). Since this initial discovery, ANCA has been found to be associated with a number of Systemic Vasculitides (SV). ANCA is now recognized to include two primary specificities: C-ANCA directed against Proteinase-3 (PR-3), and P-ANCA directed against Myeloperoxidase (MPO). Testing for both P-ANCA and C-ANCA is highly recommended in the laboratory workup of patients who present with clinical features suggestive of SV. The clinical syndromes most frequently associated with ANCA are as follows:

Wegener's granulomatosis (2) Polyarteritis (3)  
"Overlap" Vasculitis (4) Idiopathic Crescentic Glomerulonephritis (ICGN) (5) Kawasaki Disease (6)  
Although the initial identification of C-ANCA and P-ANCA was based on the indirect immunofluorescence procedures, further identification and purification of PR-3 and MPO has resulted in the development of enzyme immunoassays (ELISA) for both PR-3 and MPO.

### PRINCIPLE OF THE EUSA ASSAY

The ImmunoSpec Proteinase-3 (PR3) ELISA test system is designed to detect IgG class antibodies to PR-3 in human sera. Wells of plastic microwell strips are sensitized by passive absorption with PR3 antigen. The test procedure involves three incubation steps:

1. Test sera (properly diluted) are incubated in antigen coated microwells. Any antigen specific antibody in the sample will bind to the immobilized antigen. The plate is washed to remove unbound antibody and other serum components.
2. Peroxidase Conjugated goat anti-human IgG (y chain specific) is added to the wells and the plate is incubated. The Conjugate will react with IgG antibody immobilized on the solid phase in step 1. The wells are washed to remove unreacted Conjugate.
3. The microwells containing immobilized peroxidase Conjugate are incubated with peroxidase Substrate Solution. Hydrolysis of the Substrate by peroxidase produces a color change. After a period of time the

reaction is stopped and the color intensity of the solution is measured photometrically. The color intensity of the solution depends upon the antibody concentration in the original test sample.

### MATERIAL PROVIDED

Each kit contains the following components in sufficient quantities to perform the number of tests indicated on packaging label. Note: All reactive reagents contain sodium azide as a preservative at a concentration of 0.1% (w/v).

1. **Component list containing lot specific** information is inside the kit box.
2. **Plate.** 96 wells configured in twelve 1 x8-well strips coated with inactivated PR3 antigen. The strips are packaged in a strip holder and sealed in an envelope with desiccant.
3. **Conjugate.** Conjugated (horseradish peroxidase) goat anti-human IgG (y chain specific). Ready to use. One, 15 mL vial with a white cap.
4. **Positive Control** (Human Serum). One, 0.35 mL vial with a red cap.
5. **Calibrator** (Human Serum). One .5 mL vial with a blue cap.
6. **Negative Control** (Human Serum). One, 0.35 mL vial with a green cap.
7. **Sample Diluent.** One 30 mL bottle (green cap) containing Tween-20, bovine serum albumin and phosphate-buffered-saline, (pH 7.2 ± 0.2). Ready to use. Note: Shake Well Before Use.  
NOTE: Diluent will change color in the presence of serum.
8. **TMB:** One 15 mL amber bottle (amber cap) containing 3,3',5,5'-tetramethylbenzidine (TMB). Ready to use. Contains DMSO < 15% (w).
9. **Stop solution:** One 15 mL bottle (red cap) containing 1M<sub>2</sub>SO<sub>4</sub>, 0.7M HCl. Ready to use.
10. **Wash buffer concentrate** (10X): dilute 1 part concentrate + 9 parts deionized or distilled water. One 100 mL bottle (clear cap) containing a 10X concentrated phosphate-buffered-saline and Tween-20 solution (blue solution). NOTE: 1X solution will have a pH of 7.2 ± 0.2.
11. **Package insert** providing instructions for use.

### PRECAUTIONS

1. For In vitro Diagnostic Use.
2. Normal precautions exercised in handling laboratory reagents should be followed. In case of contact with eyes, rinse immediately with plenty of water and seek medical advice. Wear suitable protective clothing, gloves, and eye/face protection. Do not breathe vapor. Dispose of waste observing all local, state, and federal laws.
3. The wells of the ELISA plate do not contain viable organisms. However, the strips should be considered POTENTIALLY BIOHAZARDOUS MATERIALS and handled accordingly.
4. The human serum controls are POTENTIALLY BIOHAZARDOUS MATERIALS. Source materials from which these products were derived were found negative for HIV-1 antigen, HBsAg, and for antibodies against HCV and HIV by approved test methods. However, since no test method can offer complete assurance that infectious agents are absent, these products should be handled at the Biosafety Level 2 as recommended for any potentially infectious human serum or blood specimen in the Centers for Disease Control/National Institutes of Health manual "Biosafety in Microbiological and Biomedical Laboratories": current edition; and OSHA's Standard for Bloodborne Pathogens (9).
5. Adherence to the specified time and temperature of incubations is essential for accurate results. All reagents must be allowed to reach room temperature (20-25°C) before starting the assay. Return unused reagents to refrigerated temperature immediately after use.
6. Improper washing could cause false positive or false negative results. Be sure to minimize the amount of any residual wash solution; (e.g., by blotting or aspiration) before

adding Conjugate or Substrate. Do not allow the wells to dry out between incubations.

7. Sample diluent, controls, wash buffer, and conjugate contain sodium azide at a concentration of 0.1% (w/v). Sodium azide has been reported to form lead or copper azides in laboratory plumbing which may cause explosions on hammering. To prevent, rinse sink thoroughly with water after disposing of solution containing sodium azide.

8. The Stop Solution is TOXIC. Causes burns. Toxic by inhalation, in contact with skin and if swallowed. In case of accident or if you feel unwell, seek medical advice immediately.

9. The TMB Solution is HARMFUL. Irritating to eyes, respiratory system and skin.

10. The Wash Buffer concentrate is an IRRITANT. Irritating to eyes, respiratory system and skin.

11. Wipe bottom of plate free of residual liquid and/or fingerprints that can alter optical density (OD) readings.

12. Dilution or adulteration of these reagents may generate erroneous results.

13. Reagents from other sources or manufacturers should not be used.

14. TMB Solution should be colorless, very pale yellow, very pale green, or very pale blue when used. Contamination of the TMB with conjugate or other

oxidants will cause the solution to change color prematurely.

Do not use the TMB if it is noticeably blue in color. To help reduce the possibility of contamination, refer to Test Procedure, Substrate Incubation section to determine the amount of TMB to be used.

15. Never pipette by mouth. Avoid contact of reagents and patient specimens with skin and mucous membranes.

16. Avoid microbial contamination of reagents. Incorrect results may occur.

17. Cross contamination of reagents and/or samples could cause erroneous results 18. Reusable glassware must be washed and thoroughly rinsed free of all detergents.

19. Avoid splashing or generation of aerosols.

20. Do not expose reagents to strong light during storage or incubation.

21. Allowing the microwell strips and holder to equilibrate to room temperature prior to opening the protective envelope will protect the wells from condensation.

22. Wash solution should be collected in a disposal basin. Treat the waste solution with 10% household bleach (0.5% sodium hypochlorite). Avoid exposure of reagents to bleach fumes.

23. Caution: Liquid waste at acid pH should be neutralized before adding to bleach solution.

24. Do not use ELISA plate if the indicator strip on the desiccant pouch has turned from blue to pink.

25. Do not allow the conjugate to come in contact with containers or instruments that may have previously contained a solution utilizing sodium azide as a preservative. Residual amounts of sodium azide may destroy the conjugate's enzymatic activity.

26. Do not expose any of the reactive reagents to bleach-containing solutions or to any strong odors from bleach-containing solutions. Trace amounts of bleach (sodium hypochlorite) may destroy the biological activity of many of the reactive reagents within this kit.

#### MATERIALS REQUIRED BUT NOT PROVIDED:

- ELISA microwell reader capable of reading at a wavelength of 450nm.
- Pipettes capable of accurately delivering 10 to 200µL
- Multichannel pipette capable of accurately delivering (50-200 µL)
- Reagent reservoirs for multichannel pipettes.
- Wash bottle or microwell washing system.
- Distilled or deionized water.
- One liter graduated cylinder.
- Serological pipettes.
- Disposable pipette tips.
- Paper towels.
- Laboratory timer to monitor incubation steps.

- Disposal basin and disinfectant. (example: 10% household bleach, 0.5% sodium hypochlorite.)

#### STORAGE CONDITIONS

1. Store the unopened kit between 2° and 8°C.
2. Coated microwell strips: Store between 2° and 8°C. Extra strips should be immediately resealed with desiccant and returned to proper storage. Strips are stable for 60 days after the envelope has been opened and properly resealed and the indicator strip on the desiccant pouch remains blue.
3. Conjugate: Store between 2° and 8°C. DO NOT FREEZE.
4. Calibrator, Positive Control and Negative Control: Store between 2° and 8°C.
5. TMB: Store between 2° and 8°C.
6. Wash Buffer concentrate (10X): Store between 2° and 25°C. Diluted wash buffer (1X) is stable at room temperature (20° and 25°C) for up to 7 days or for 30 days between 2° and 8°C.
7. Sample Diluent: Store between 2° and 8°C.
8. Stop Solution: Store between 2° and 8°C.

#### SPECIMEN COLLECTION

1. It is recommended that specimen collection be carried out in accordance with NCCLS document M29: Protection of Laboratory Workers from Infectious Disease.
2. No known test method can offer complete assurance that human blood samples will not transmit infection. Therefore, all blood derivatives should be considered potentially infectious.
3. Only freshly drawn and properly refrigerated sera obtained by approved aseptic venipuncture procedures should be used in this assay (7, 8). No anticoagulants or preservatives should be added. Avoid using hemolyzed, lipemic, or bacterially contaminated sera.
4. Store sample at room temperature for no longer than 8 hours. If testing is not performed within 8 hours, sera may be stored between 2° and 8°C for no longer than 48 hours. If delay in testing is anticipated, store test sera at -20°C or lower. Avoid multiple freeze/thaw cycles that may cause loss of antibody activity and give erroneous results.

#### GENERAL PROCEDURE

- 1 Remove the individual components from storage and allow them to warm to room temperature (20° and 25°C).
2. Determine the number of microwells needed. Allow six Control/Calibrator determinations (one Blank, one Negative Control, three Calibrators and one Positive Control) per run. A Reagent Blank should be run on each assay. Check software and reader requirements for the correct Controls/Calibrator configurations. Return unused strips to the resealable pouch with desiccant, seal, and return to storage between 2° and 8°C.

EXAMPLE PLATE SET-UP		
	1	2
A	Blank	Patient 3
B	Neg. Control	Patient 4
C	Calibrator	Etc.
D	Calibrator	
E	Calibrator	
F	Pos. Control	
G	Patient 1	
H	Patient 2	

3. Prepare a 1:21 dilution (e.g.: 10µL of serum + 200µL of Sample Diluent. NOTE: Shake Well Before Use) of the Negative Control, Calibrator, Positive Control, and each patient serum, The Sample Diluent will undergo a color change confirming that the specimen has been combined with the diluent.
  4. To individual wells, add 100µL of each diluted control, calibrator and sample. Ensure that the samples are properly mixed. Use a different pipette tip for each sample.
  5. Add 100µL of Sample Diluent to well A1 as a reagent blank. Check software and reader requirements for the correct reagent blank well configuration.
  6. Incubate the plate at room temperature (20° and 25°C) for 25± 5 minutes.
  7. Wash the microwell strips 5x
- A. Manual Wash Procedure:
- a. Vigorously shake out the liquid from the wells.

- b. Fill each microwell with Wash Buffer. Make sure no air bubbles are trapped in the wells.
- c. Repeat steps a. and b. for a total of 5 washes.
- d. Shake out the wash solution from all the wells. Invert the plate over a paper towel and tap firmly to remove any residual wash solution from the wells. Visually inspect the plate to ensure that no residual wash solution remains. Collect wash solution in a disposable basin and treat with 0.5% sodium hypochlorite (bleach) at the end of the days run.

**B. Automated Wash Procedure:**

If using an automated microwell wash system, set the dispensing volume to 300-350µL lwell. Set the wash cycle for 5 washes with no delay between washes. If necessary, the microwell plate may be removed from the washer, inverted over a paper towel and tapped firmly to remove any residual wash solution from the microwells.

8. Add 100µL of the Conjugate to each well, including reagent blank well, at the same rate and in the same order as the specimens were added.
9. Incubate the plate at room temperature (20° and 25°C) for 25 ± 5 minutes
10. Wash the microwells by following the procedure as described in step 7.
11. Add 100µL of TMB to each well, including reagent blank well, at the same rate and in the same order as the specimens were added.
12. Incubate the plate at room temperature (20° and 25°C) for 10 to 15 minutes.
13. Stop the reaction by adding 50µL of Stop Solution to each well, including reagent blank well, at the same rate and in the same order as the TMB was added. Positive samples will turn from blue to yellow. After adding the Stop Solution, tap the plate several times to ensure that the samples are thoroughly mixed.
14. Set the microwell reader to read at a wavelength of 450nm and measure the optical density (OD) of each well against the reagent blank. The plate should be read within 30 minutes after the addition of the Stop Solution.

**QUALITY CONTROL**

1. Each time the assay is run the Calibrator must be run in triplicate. A reagent blank, Negative Control, and Positive Control must also be included in each assay.
2. Calculate the mean of the three Calibrator wells. If any of the three values differ by more than 15% from the mean, discard that value and calculate the mean using the remaining two wells.
3. The mean OD value for the Calibrator and the OD values for the Positive and Negative Controls should fall within the following ranges:

	<u>UU Range</u>
Negative Control	≤ 0.250
Calibrator	≥ 0.300
Positive Control	≥ 0.500

- a. The OD of the Negative Control divided by the mean OD of the Calibrator should be ≤ 0.9.
- b. The OD of the Positive Control divided by the mean OD of the Calibrator should be ≥ 1.25.
- c. If the above conditions are not met the test should be considered invalid and should be repeated.

4. The Positive Control and Negative Control are intended to monitor for substantial reagent failure and will not ensure precision at the assay cut-off.
5. Additional controls may be tested according to guidelines or requirements of local, state, and/or federal regulations or accrediting organizations.
6. Refer to NCCLS document C24: Statistical Quality Control for Quantitative Measurements for guidance on appropriate QC practices.

**INTERPRETATION OF RESULTS**

**A. Calculations:**

**1. Correction Factor**

A cutoff OD value for positive samples has been determined by the manufacturer and correlated to the Calibrator. The correction factor (CF) will allow you to determine the cutoff value for positive samples and to correct for slight day-to-day variations in test results. The correction factor is determined for each lot of kit components and is printed on the Component list located in the kit box.

**2. Cutoff OD Value**

To obtain the cutoff OD value, multiply the CF by the mean OD of the Calibrator determined above. (CF x mean OD of Calibrator = cutoff OD value)

**3. Index Values or OD Ratios**

Calculate the Index Value or OD Ratio for each specimen by dividing its OD value by the cutoff OD from step 2.

Example:

Mean OD of Calibrator	=	0.793
Correction Factor (CF)	=	0.25
Cut off OD	=	0.793 x 0.25 = 0.198
Unknown Specimen OD	=	0.432
Specimen Index Value or OD Ratio	=	0.432 / 0.198 = 2.18

**B. Interpretations:**

Index Values or CD ratios are interpreted as follows:

	<u>Index Value or OD Ratio</u>
Negative Specimens	≤ 0.90
Equivocal Specimens	0.91 to 1.09
Positive Specimens	≥ 1.10

**LIMITATION OF THE ASSAY**

1. A diagnosis should not be made on the basis of anti-PR3 ELISA results alone. Test results for anti-PR3 should be interpreted in conjunction with the clinical evaluation and the results of other diagnostic procedures.
2. The performance characteristics of this device have not been established for lipemic, hemolyzed and icteric specimens; therefore, these specimens should not be tested with this assay.
3. The results of this assay are not diagnostic proof of the presence or absence of disease. Immunosuppressive therapy should not be started based on a positive result.

**EXPECTED RESULTS**

In a study conducted by Immunospec, 90 normal donor sera from Southwestern United States were evaluated for Proteinase-3 autoantibodies. Cf the 90 tested, none were positive. In another study using 113 specimens which were sent to a reference laboratory in Northeastern United States, eight (8/113 = 7.1%) were positive for anti-Proteinase-3 IgG. Taken together, these studies demonstrate that the incidence of IgG antibody to Proteinase-3 is relatively rare.

**PERFORMANCE CHARACTERISTICS**

**Comparative Study:**

An in-house comparative study was performed to demonstrate the equivalence of the Immunospec Proteinase-3 IgG ELISA test system to another commercially available PR3 IgG ELISA test system. Performance was evaluated using 315 specimens; 196 disease-state specimens, 113 specimens which were sent to a reference laboratory in the Northeastern United States for routine ANCA serology, and 6 specimens which were previously tested and found to be reactive for ANCA. A summary of the clinical specimens appears in Table 1 below. The results of the investigation have been summarized in Table 2 below.

Table 1. Summary of Clinical Specimens

n	Male	Female	AGE			Comments
			High	Low	Mean	
45	18	27	82	14	54.7	Disease Category: Wegener's Granulomatosis
41	21	20	100	22	63.2	Disease Category: Idiopathic Necrotizing and Crescentic Glomerulonephritis
41	16	25	87	20	63.1	Disease Category: Microscopic Polyarteritis
39	17	22	94	11	60.8	Disease Category: Pulmonary Renal Syndrome
30	15	15	78	3	43.4	Vasculitis/Glomerulonephritis Disease Controls, Non-ANCA related vasculitis.
6	Information Not Available					Previously tested ANCA positive, no diagnosis available
113	Information Not Available					Specimens sent to a reference laboratory for routine ANCA serology

Table 2: Calculation of Relative Sensitivity, Specificity, and Agreement  
Immunospec PR3 Results

		+	-	±*	Totals
Commercial PR3 ELISA	+	56	6	0	62
	-	3	211	4	218
Test System	±*	11	23	1	35
	Totals	70	240	5	315

\*Equivocal specimens were excluded from all calculations.

Relative Sensitivity = 56/62 = 90.3%      95% Confidence Interval\*\* = 83 to 97.7%

Relative Specificity = 211/214 = 98.6%      95% Confidence Interval\*\* = 97.0 to 100%

Relative Agreement = 267/276 = 96.7%      95% Confidence Interval\*\* = 94.6 to 98.8%

\*\*95% confidence intervals calculated using the exact method.

### Reproducibility:

To evaluate both intra-assay and inter-assay reproducibility, six specimens were tested; eight replicates each, on each of three days. These results were then used to calculate mean unit values, standard deviations, and percent CV. Two of the specimens were strong positives, two were clearly negative, and two were near the assay cut off. The results of the study have been summarized below.

**Table 3. Immunospes PR3 IgG ELISA; Results of Precision Testing**

Day 1: Intra-Assay Reproducibility.

Specimen	Mean Result (AAU/mL)	Standard Deviation	Percent CV	Replicates
1	1077	39	3.6	8
2	468	12	2.6	8
3	38	2	6.6	8
4	195	6	3.3	8
5	26	2	8.5	8
6	1687	61	3.6	8

Day 2: Intra-Assay Reproducibility.

Specimen	Mean Result (AAU/mL)	Standard Deviation	Percent CV	Replicates
1	967	60	6.2	8
2	406	27	6.6	8
3	42	3	8.1	8
4	192	8	4.1	8
5	35	2	6.5	8
6	1702	74	4.3	8

Day 3: Intra-Assay Reproducibility.

Specimen	Mean Result (AAU/mL)	Standard Deviation	Percent CV	Replicates
1	1085	24	2.2	8
2	423	20	4.6	8
3	36	4	9.8	8
4	178	13	7.6	8
5	28	5	17.2	8
6	1628	54	3.3	8

Inter-Assay Reproducibility; All Days Combined:

Specimen	Mean Result (AAU/mL)	Standard Deviation	Percent CV	Replicates
1	1043	69	6.6	24
2	432	33	7.7	24
3	39	4	10.4	24
4	189	12	6.5	24
5	30	5	15.4	24
6	1672	69	4.1	24

### Cross Reactivity:

To evaluate the test system for potential cross-reactivity to other autoantibodies, eight specimens which were positive for antibodies to nuclear antigens (ANA) on HEp-2 cells were tested. Two of the specimens demonstrated a homogenous pattern, two demonstrated a nucleolar pattern, two demonstrated the centromere pattern, and two demonstrated a speckled pattern. The results of this study have been summarized in Table 4 below. The results of this investigation indicate that cross reactivity with other antinuclear antibodies is not likely.

Table 4. Results of the Cross Reactivity Investigation.

Sample Number	ANA HEp-2 IFA RESULTS		PR3 IgG ELISA RESULTS	
	Pattern	Endpoint Titer	Optical Density	AAU/mL
1	Homogeneous	1:1280	0.056	53
2	Homogeneous	1:640	0.014	13
3	Speckled	1:2560	0.030	28
4	Nucleolar	1:1280	0.069	66
5	Centromere	1:1280	0.060	58
6	Centromere	1:1280	0.026	25
7	Speckled	1:5120	0.049	48
8	Nucleolar	1:10240	0.019	48

### REFERENCES

- Davies D, Moran ME, 85:606, 1982
- van der Woude FJ, Rasmussen N, Lobatto 5, Wiik A, Permin H, van Es LA: Autoantibodies against neutrophils and monocytes: Tools for diagnosis and marker of disease activity in Wegener's granulomatosis. Lancet 1:425-429, 1985.
- Feehally J, Wheeler DC, Walls J, et al: A case of microscopic polyarteritis associated with antineutrophilia cytoplasmic antibodies. Clin. Nephrol. 27: 214-215, 1987
- Falk RJ, Becker M, Terrell R, Jennette JC: Antigen specificity of P-ANCA and C-ANCA (Abstract). The 3rd International Workshop on ANCA, Washington, DC 1990: 2-3.
- Jennette JC, Falk RJ: Anti-neutrophil cytoplasmic autoantibodies with specificity for myeloperoxidase in patients with systemic vasculitis and idiopathic necrotizing and crescentic glomerulonephritis. N.Engl.J. Med. 318: 1651-1657, 1988.
- Savage COS, Tizard J, Jayne D, et al: Anti-neutrophil cytoplasmic antibodies in Kawasaki disease. Arch. Dis. Child. 64: 360-363, 1989.
- Procedures for the collection of diagnostic blood specimens by venipuncture. Second Edition: Approved Standard (1984). Published by National Committee for Clinical Laboratory Standards.
- Procedures for the Handling and Processing of Blood Specimens. NCCLS Document H18-A, Vol. 10, No. 12, Approved Guideline, 1990.
- U.S. Department of Labor, Occupational Safety and Health Administration: Occupational Exposure to Bloodborne Pathogens, Final Rule. Fed. Register 56:64175-64182, 1991.

### **ABBREVIATED TEST PROCEDURE**

1. Dilute Serum 1:21
2. Add diluted serum to microwell 100  $\mu$ L/well
3.  $\longrightarrow$  *Incubate 20 to 30 minutes*
4. Wash
5. Add Conjugate – 100  $\mu$ L/well
6.  $\longrightarrow$  *Incubate 20 to 30 minutes*
7. Wash
8. Add TMB 100  $\mu$ L/well
9.  $\longrightarrow$  *Incubate 10 to 15 minutes*
10. Add Stop Solution 50  $\mu$ L/well - Mix
11. READ



**European Authorized Representative:**

CEpartner4U , Esdoornlaan 13, 3951DB Maarn  
. The Netherlands. Tel.: +31 (0)6.516.536.26



**Manufacturer:**

IMMUNOSPEC CORPORATION  
7018 Owensmouth Ave. Suite 103  
Canoga Park, CA, 91303  
Phone: 818-710-1281  
Fax: 818-936-0121

PR 3  
REVISION DATE: 09/23/05  
R2028E